

# NeuroBlate® System and Atama Instructions for Use Compatible MRI Systems Supplement

## Contents

1.	Overview .....	1
1.1.	Warning and Identification Labels .....	1
2.	NeuroBlate Compatible MRI Systems.....	2

Patent Notice under 35 U.S.C. §287(a): Monteris Medical products and associated processes are subject to the United States patents and pending applications.

NeuroBlate, Monteris Medical, SideFire, FullFire, Atama, AutoLITT, Monteris Mini-Bolt, the M Logo, the N Logo, the A Logo and Fusion-S Pro are trademarks of Monteris Medical. ©2023 Monteris. All Rights Reserved. ™ and ® denote Trademarks and Registered Trademarks of Monteris Medical Corporation.

## 1. Overview





This document describes the MRI Systems that are compatible with the NeuroBlate System.

This document supplements information for the following NeuroBlate System IFUs:

- > 80554 – NeuroBlate System v3.17 SW
- > 80512 – NeuroBlate System v3.16 SW
- > 80327 – Atama System

### 1.1. WARNING AND IDENTIFICATION LABELS

Symbols displayed on Monteris products or in their documentation are:

	<p><b>MR Unsafe</b> - item is <b>NOT</b> MRI compatible and is known to pose a hazard in MR environments. This equipment should not be taken into the MRI room within the 5 Gauss perimeter line.</p>
	<p><b>MR Conditional</b> - the item has been demonstrated to pose no known hazards in a 1.5T or 3.0T MR environments as described in Section 4.4 of the Instructions for Use.</p>
	<p><b>MR Safe</b> - the item poses <b>NO</b> known hazards in ALL MR environments.</p>
	<p>Refer to instructions</p>

## 2. NeuroBlate Compatible MRI Systems

The NeuroBlate System and Atama is compatible with only the following 1.5T or 3.0T MRI systems.



1.5 T / 3.0 T

- > Siemens - minimum software versions VB17, VB19, VD13, VE11, (VXA20, VXA30, VXA50, and VXA51 with Access-I enabled systems)
  - Symphony (1.5T)
  - Avanto (1.5T)
  - Trio (3.0T)
  - Espree (1.5T)
  - Verio (3.0T)
  - Aera (1.5T)
  - Skyra (3.0T)
  - Prisma (3.0T)
  - Sola (1.5T)
  - Vida (3.0T)
- > IMRIS NEURO II (1.5T) or NEURO III (3.0T) - Siemens MRI software versions listed above
- > GE - minimum software version 12.X, 15.X to 30.X
  - Signa HDx/HDxT (1.5T & 3.0T)
  - Optima MR450/450W (1.5T)
  - Discovery MR750/750W (3.0T)
  - Signa Artist (1.5T)
  - Signa Architect (3.0T)
  - Signa Premier (3.0T)
- > Philips – software version R3, R4, R5
  - Achieva (1.5T & 3.0T)
  - Ingenia (1.5T and 3.0T)
- > Intera (1.5T) software version v3



### U.S. Office

131 Cheshire Lane, Suite 100  
Minnetonka, MN 55305, USA  
+1.866.799.7655

### Canadian Office

Unit 1B-25 Scurfield Blvd.  
Winnipeg, MB R3Y1G4, Canada  
+1.204.272.2220

